



UKCA – Declaration of Conformity

We declare that the product **Hand Held Instrument**

series **MH 3111**

placed on the market by **SIKA Dr. Siebert & Kühn GmbH & Co. KG**

comply with

Electromagnetic Compatibility Regulations 2016

Electrical Equipment (Safety) Regulations 2016

The Restriction of the Use of Certain Hazardous Substances in
Electrical and Electronic Equipment Regulations 2012

The devices comply with following technical specifications

| | |
|--------------------------------|--|
| BS EN 61326-1:2013-02 | Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements |
| BS EN 61010-1:2010+ A1:2019 | Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements |
| BS EN IEC 63000:2018 | Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances |

Kaufungen, 07.06.2021

i.V.

Dipl.- Ing. K. Ulloth
(Manager CE- Coordination and Safety Supervisor)
(CE- Koordinator und Produktsicherheitsbeauftragter)